

P U L S A R 250 - ES00005-A - P R O S E R I E S



P U L S A R 250

T H E A D M I R A L



REF: ES00005-A

UDI-ID: 80536089401262

Up to 250 tests/hour



P R O S E R I E S :



C O R E

175
Tests/hour



O R I O N 240

240
Tests/hour



P U L S A R 250

250
Tests/hour



O R I O N 320

320
Tests/hour



P U L S A R 250



REF: ES00005-A

UDI-ID: 80536089401262

Up to 250 tests/hour



Physical characteristics		
SIZE / WEIGHT	Height: 54 cm (21.3 in) Depth: 77 cm (30.3 in) Width: 63 cm (24.8 in)	Weight: 47 kg (103,6 lbs)
POWER SUPPLY	100/ 240 Vac, 50 / 60 Hrz, single phase with ground Independent on-off switch for refrigerated reagent plate Fuse compartment / fuses: 2.5 Amp @ 230 Vac, 4 Amp @ 115 Vac Power consumption: less than 200 VA (external PC excluded) Ground resistance: less than 0.1 Ohm Leakage current: less than 2.5 mA	
SAMPLING ARM	1 sampling needle, 110 mm needle stroke Capacitive liquid level detector 3-axes-needle shock sensor	
DILUTER SYRINGE	Long life plunger Syringe capacity, 255 µL Syringe resolution, 0.05 µL	
HYDRAULIC SYSTEM	8 self-priming peristaltic pumps with replaceable neoprene tube cassette (life 500 hrs) 2 vacuum pumps 3-way Pinch valve Manifold Containers*: Water, 20L; Cleaning solution, 2L; Waste, 20L * equipped with level sensor and safety connections Water consumption: less than 1 L/h ±10%	
WASH STATION	Needles: 4 dispensing, 5 aspiration, 1 cleaning (8 step washing sequence for each cuvette) cleaning solution on needle 2.	
REAGENTS AND SAMPLE ROTOR	Removable rotor Samples and reagents barcode reader 36 bottles, 50 mL or 20 mL /optional 38mL or 15mL (up to 1440 ml total) Reagents barcode reader Samples tray, 36+36 numbered positions, 36 barcoded tubes of 12 -13 mm, 5 - 7 mL / 36 screw-cups of 2 mL/ free standard /control/ urgent positions	
CUVETTE ROTOR REACTION CELLS	80 washable direct reading cuvettes which allow up to 30 000 tests per rotor Optical path 6 mm, 210 - 350 µL reaction volume 100W heating resistance, temperature sensor, safety thermostat	
OPTICAL GROUP	1 halogen lamp (6 V, 10 W) with extended UV emission 2 focusing lenses, optical glass 10-position filter disk: 8 positions provided with interference filters of 340, 405, 505, 546, 578, 600, 650, 700 nm wavelengths, 1 free position and 1 solid position for dark reading ±2 nm on peak wavelength, band pass of ±10 nm	
PHOTOAMPLIFIER	Photoelectric detector Signal amplifier Response range, 340 nm to 900 nm Photometric range, 0 to 3 Abs Linearity, ±0.5% from 0.1 to 1.5 Abs Precision: 1% CV or 1 mAbs min. (0.1 to 1.5 Abs) Stability: daily reader offset, less than 1% drift per day	
CONTROL	Real-time multitasking microprocessor based control Easy access to the electronics	
EXTERNAL COMPUTER	(Minimum requirements for Software  v.1.7) CPU: Intel i3 or superior RAM: 8GB SSD: 100GB Hard Disk I/O: USB 2.0 port Display: minimum resolution 1280x768 OS: Microsoft Windows® 10 or 11 Framework: .NET framework 4.6	

Operation features	
PIPETTING	Volume: sample, 1 - 255 µl; reagent, 1-255 µL Precision: 2 CV% at 1 µl; 1.5 CV% at 2 µl; 1 CV% at 4 µL Mixing by sample needle upon dispensation
REACTION	Reaction volume, 100 - 500 µL
SAMPLE DILUTION	In-needle dilution if allowed by method's sample volumes Automatic pre-dilution in a reaction cuvette, up to 1:100
TEMPERATURE	Reagents and samples refrigeration, about 15 °C below room temperature
CONTROL	Reaction cells, heating unit can be set from room temperature up to 42 °C ±0.2 °C (108 °F ±0.5 °F)
TYPES OF TESTS	Endpoint, Bichromatic endpoint, Differential endpoint, Differential endpoint sample blank, Fixed Time, Kinetic, Kinetic bichromatic
TEST RUNS	Random / Urgent (sample attribute)
MEASUREMENT RATES	250 tests/hour in single reagent mode Maximum incubation + reading time: 800 seconds Typical precision, endpoint 2.0 CV% / kinetic 2.0 CV%
START-UP	The start-up procedure is run daily: self-test, reader offset of optics, wash and check of all cuvettes
CALIBRATION	Reagent blank subtraction, 1 to 8 standards per test method Linear: factor, linear, linear regression Non linear (5 interpolation types): cubic-spline, poly- linear, multi-parameter, logit-log four parameters and five parameters Free standard /control positions (5 mL tubes or 1 mL cups) Results can be recalculated when changing factor or calibration curve
MAINTENANCE	Procedures programmed by component life counters
PRINTING REPORTS	Single test, complete sample, work sheet, method and QCs Automatic sample reports upon test completion if requested
NEEDLE WASHING	Sampling needle washed internally and externally with system solution after every operation

Connections	
POWER	Standard VDE removable power cord
EXTERNAL PC	USB port
HOST/ LIS	Ethernet LAN (samples, work list, results) Standard ASTM ASCII protocol
Database	
WORKLIST/ SAMPLES	For each worklist: unlimited number of samples, unlimited number of tests, up to 99 sheets of tests per worklist. Tests archive with powerful search tools Patient management
TEST METHODS	Unlimited number of methods in PC memory 100 active methods
QUALITY CONTROL	Three-level controls per test, one month monitoring Reagent/calibrator/control lot monitoring, Exclusion of failing results from graphic and statistics
ERROR LOG	Automatically stored at run-time, can be viewed or printed Powerful on-line monitoring